

Certificate



Certificate no.

CU 72120083 01

License Holder:

Slaughter Company, Inc.
28105 N. Keith Drive
Lake Forest IL 60045-4546
USA

Manufacturing Plant:

Extech Electronics Co., Ltd.
4F, No. 252, Nan Yang Street
221 Xizhi District, New Taipei City
Taiwan

Test report no.: USA- 10033741 001

Client Reference:

Tested to: UL 61010-1:2004 R10.08
CAN/CSA-C22.2 NO. 61010-1-04

Certified Product: 500VA ACW/DCW/IR/GB Tester

License Fee - Units

Model Designation:	4520	7
Rated Voltage:	AC 115V/230V; 50/60Hz	
Rated Current:	15A/10A max.	
High voltage Output Ratings:	AC 5000V max., 100mA; DC 6000V max., 10mA	
Protection Class:	I	
Installation Category:	II	

Appendix: 1, 1-12

7

Licensed Test mark:



Signature

Dipl.-Ing./F. Staelzel
QA Certification Officer

Date of Issue
(day/mo/yr)
12/01/2012